

## Surface Energy Balance System (SEBS) Instrument Handbook

DR Cook  
RC Sullivan

Updated February 2024



## **DISCLAIMER**

This report was prepared as an account of work sponsored by the U.S. Government. Neither the United States nor any agency thereof, nor any of their employees, makes any warranty, express or implied, or assumes any legal liability or responsibility for the accuracy, completeness, or usefulness of any information, apparatus, product, or process disclosed, or represents that its use would not infringe privately owned rights. Reference herein to any specific commercial product, process, or service by trade name, trademark, manufacturer, or otherwise, does not necessarily constitute or imply its endorsement, recommendation, or favoring by the U.S. Government or any agency thereof. The views and opinions of authors expressed herein do not necessarily state or reflect those of the U.S. Government or any agency thereof.

# Surface Energy Balance System (SEBS) Instrument Handbook

DR Cook  
RC Sullivan  
Both at Argonne National Laboratory

Updated February 2024

How to cite this document:

Cook, DR, and R Sullivan. 2024. Surface Energy Balance System (SEBS) Instrument Handbook. U.S. Department of Energy, Atmospheric Radiation Measurement user facility, Richland, Washington. DOE/SC-ARM-TR-092.

Work supported by the U.S. Department of Energy,  
Office of Science, Office of Biological and Environmental Research

## Acronyms and Abbreviations

ARM	Atmospheric Radiation Measurement
DQ	Data Quality
DQ Explorer	Data Quality Explorer
DQR	Data Quality Report
ECOR	Eddy Correlation Flux Measurement System
EF	extended facility
EFSCO	EF Surface Conditions Observations
ENA	Eastern North Atlantic
IMMS	Instrument Mentor Monthly Summaries
IR	infrared
MET	surface meteorological instrumentation
netCDF	Network Common Data Form
NSA	North Slope of Alaska
OLI	Oliktok Point, Alaska
PRTD	platinum resistance temperature detector
QC	quality control
REBS	Radiation and Energy Balance Systems, Inc.
RMS	root mean square
SEBS	surface energy balance system
SGP	Southern Great Plains
TWP	Tropical Western Pacific
VAP	value-added product
WMO	World Meteorological Organization
WPL	Webb, Pearman, Leuning

# Contents

Acronyms and Abbreviations .....	iii
1.0 General Overview .....	1
2.0 Contacts .....	1
2.1 Mentor .....	1
2.2 Instrument Developer .....	1
3.0 Deployment Locations and History .....	2
4.0 Data Descriptions and Examples .....	2
4.1 Data File Contents .....	2
4.1.1 Primary Variables and Expected Uncertainty .....	2
4.1.2 Secondary/Underlying Variables .....	3
4.1.3 Diagnostic Variables .....	3
4.1.4 Data Quality Flags.....	3
4.1.5 Dimension Variables .....	4
4.2 Annotated Examples .....	4
4.3 User Notes and Known Problems .....	4
4.4 Frequently Asked Questions .....	4
5.0 Data Quality.....	4
5.1 Data Reviews by Instrument Mentor.....	4
5.2 Data Assessments by Site Scientist/Data Quality Office .....	5
6.0 Instrument Details.....	7
6.1 Detailed Description.....	7
6.1.1 List of Components .....	7
6.1.2 System Configuration and Measurement Methods .....	7
6.1.3 Specifications .....	8
6.2 Theory of Operation.....	8
6.3 Calibration.....	8
6.3.1 Theory .....	8
6.3.2 Procedures.....	8
6.3.3 History.....	8
6.4 Operation and Maintenance .....	8
6.4.1 User Manual .....	8
6.4.2 Additional Documentation .....	9
7.0 Glossary.....	9
8.0 Citable References.....	9

## Tables

1 SEBS status. .... 2

## 1.0 General Overview

A surface energy balance system (SEBS) has been installed collocated with each deployed eddy correlation flux measurement system (ECOR) at the U.S. Department of Energy (DOE) Atmospheric Radiation Measurement (ARM) user facility Southern Great Plains (SGP) observatory, North Slope of Alaska (NSA) observatory, Eastern North Atlantic (ENA) observatory, and the first, second, and third ARM Mobile Facility (AMF1, AMF2, AMF3). A SEBS was also deployed at the Tropical Western Pacific (TWP) site, before it was decommissioned. Data from these sites, including the retired TWP, are available in the ARM Data Center. The SEBS consists of upwelling and downwelling solar and infrared radiometers within one net radiometer, a wetness sensor, and soil measurements. The SEBS measurements allow the comparison of ECOR sensible and latent heat fluxes with the energy balance determined from the SEBS and provide information on wetting of the sensors for data quality purposes.

## 2.0 Contacts

### 2.1 Mentor

Ryan C. Sullivan  
Environmental Science Division  
Argonne National Laboratory, Bldg. 240  
9700 S. Cass Avenue  
Lemont, Illinois 60439  
Phone: 630-252-1270  
[rcsullivan@anl.gov](mailto:rcsullivan@anl.gov)

### 2.2 Instrument Developer

David R. Cook  
Environmental Science Division  
Argonne National Laboratory, Bldg. 240  
9700 S. Cass Avenue  
Lemont, Illinois 60439  
Phone: 630-252-5840  
Fax: 630-252-2959  
[drcook@anl.gov](mailto:drcook@anl.gov)

### 3.0 Deployment Locations and History

Table 1. SEBS status.

Facility	Location	Date Installed	Status
SGP/EF14	Lamont, OK	10/18/2010	Operational
SGP/EF21	Okmulgee, OK	10/20/2010	Removed 05/02/2019
SGP/EF31	Anthony, KS	11/03/2011	Removed 09/21/2012
SGP/EF33	Newkirk, OK	11/03/2011	Operational
SGP/EF37	Waukomis, OK	11/04/2011	Operational
SGP/EF38	Omega, OK	11/02/2011	Removed 06/07/2021
SGP/EF39	Morrison, OK	09/29/2015	Operational
SGP/EF41	Peckham, OK	04/26/2016	Removed 08/02/2023
ENA	Azores, Portugal	07/23/2014	Operational
AMF1	Various	01/02/2011	Operational
AMF2	Various	11/01/2010	Operational
AMF3	Oliktok, AK	06/25/2014	Removed 06/15/2021
TWP/E30	East ARM	08/11/2010	Removed 01/10/2015
TWP/E31	Beatrice Hill	08/13/2010	Removed 01/03/2015
TWP/E32	Berrimah	08/02/2010	Removed 01/10/2015
NSA/E10	Barrow	09/13/2011	Operational
NSA/E11	Point Barrow	06/03/2012	Removed 12/06/2016

### 4.0 Data Descriptions and Examples

#### 4.1 Data File Contents

##### 4.1.1 Primary Variables and Expected Uncertainty

30 minutes:

surface\_energy\_balance: 5% uncertainty

surface\_soil\_heat\_flux\_avg: 10% uncertainty

net radiation: 5% uncertainty

wetness: 5% uncertainty

##### Definition of Uncertainty

We define uncertainty as the range of probable maximum deviation of a measured value from the true value within a 95% confidence interval. Given a bias (mean) error  $B$  and uncorrelated random errors characterized by a variance  $\sigma^2$ , the root-mean-square error (RMSE) is defined as the vector sum of these,

$$RMSE = (B^2 + \sigma^2)^{1/2}.$$

( $B$  may be generalized to be the sum of the various contributors to the bias and  $\sigma^2$  the sum of the variances of the contributors to the random errors). To determine the 95% confidence interval we use the



Student's  $t$  distribution:  $t_{n;0.025} \approx 2$ , assuming the RMSE was computed for a reasonably large ensemble. Then the *uncertainty* is calculated as twice the RMSE.

#### 4.1.2 Secondary/Underlying Variables

30 minutes:

time\_offset

time

down\_short\_hemisp

up\_short\_hemisp

down\_long

up\_long

surface\_soil\_heat\_flux\_1

surface\_soil\_heat\_flux\_2

surface\_soil\_heat\_flux\_3

soil\_moisture\_1

soil\_moisture\_2

soil\_moisture\_3

soil\_temp\_1

soil\_temp\_2

soil\_temp\_3

soil\_heat\_flow\_1

soil\_heat\_flow\_2

soil\_heat\_flow\_3

corr\_soil\_heat\_flow\_1

corr\_soil\_heat\_flow\_2

corr\_soil\_heat\_flow\_3

soil\_heat\_capacity\_1

soil\_heat\_capacity\_2

soil\_heat\_capacity\_3

energy\_storage\_change\_1

energy\_storage\_change\_2

energy\_storage\_change\_3

albedo

temp\_net\_radiometer

DOY

gmthour

#### 4.1.3 Diagnostic Variables

battery\_voltage

#### 4.1.4 Data Quality Flags

The b1 data file contains basic data quality flags for most important variables; the flags indicate the variable status (bit values), as follows:

- 0x0 = value is within the specified range
- 0x1 = value is equal to “missing\_value”
- 0x2 = value is less than “valid\_min”
- 0x4 = value is greater than “valid\_max”
- 0x8 = value failed the “valid\_delta” check.

#### **4.1.5 Dimension Variables**

The global attributes section of the netCDF data file contains geographic coordinates (location) of the SEBS system and the altitude of the ground where the instrument is deployed; the “sensor location” parameter refers to the height of the instrument above the ground. The time variables denote the end of the 30-minute measuring period.

The sign convention for radiometer, net radiation, change of energy storage, and soil heat flux measurements is: positive towards the soil surface and negative away from the surface.

The net radiometer is pointed due south.

## **4.2 Annotated Examples**

Not available at this time.

## **4.3 User Notes and Known Problems**

The upwelling and downwelling solar radiation measurements often exhibit small negative values at night; these values are not physically real.

The wetness values go below the lower limit when icing occurs, and spikes above the upper limit after the ice melts. The condition most frequently occurs at NSA and Oliktok Point (OLI), but can occur at other locations where icing, snow, and frost occur (including SGP).

## **4.4 Frequently Asked Questions**

Where do I get more information about SEBS systems?

Contact the instrument mentor at [rcsullivan@anl.gov](mailto:rcsullivan@anl.gov).

## **5.0 Data Quality**

### **5.1 Data Reviews by Instrument Mentor**

- Visual QC frequency: Daily to weekly

- QC delay: Typically 1–3 days
- QC type: Instrument mentor routinely views graphic displays that include plots (day courses) of all calculated quantities and comparison plots (time series or scatter plots) of relevant parameters with data from collocated ECOR, SEBS, EBBR (SGP CF and EF39 only), and surface meteorological instrumentation (MET) (Cook et al. 2006).

Monthly reviews of the SEBS data were prepared by the mentor and submitted to the Instrument Mentor Monthly Summary (IMMS) report database until late 2014.

Data Quality Reports (DQRs) are not written for missing data or for situations when QC flags clearly show that the data are incorrect (this is true for most of the conditions listed below). DQRs are written for periods when data are incorrect, when the situation is not represented by QC flags in the data, and it is not obvious that the data should have been flagged as incorrect.

## 5.2 Data Assessments by Site Scientist/Data Quality Office

The following guidance has been provided by the SEBS mentor for use by the Data Quality Office in preparing their weekly assessment report for the SEBS systems.

SEBS Data Quality Guidance

David R. Cook

1/10/2011

**Introduction:** The best way to tell someone what to look for in assessing the SEBS data is to describe conditions that reflect correct and incorrect data. For the most part, the QC checks provide adequate guidance. However, there are conditions for which the QC flags may not provide the needed guidance to be able to interpret the correctness of the data. Therefore, please use the information below as further guidance.

**Primary Measurements:** surface\_energy\_balance, surface\_soil\_heat\_flux\_avg, and net radiation are calculated quantities, whereas wetness is a direct measurement. The QC limits set in the ingest are appropriate for the measurements (primary and otherwise), although there are times when legitimate values fall outside the QC limits (particularly for upwelling and downwelling solar radiation).

**Nuisance QC Flags:** The upwelling and downwelling minimum solar radiation flags are frequently tripped at night, exhibiting small negative values that are not physically real; these are instrument offsets.

**Comparison of Data at Adjacent SEBS Sites:** Only some measurements can be favorably compared with those at adjacent sites, keeping in mind that climate conditions from one side of the ARM site to another can differ sharply. Particular caution must be used in making comparisons between two adjacent sites when they see different vegetation surfaces (Cook et al. 2006). The upwelling radiometer and surface soil heat flux measurements from two sites may not be similar unless the ground is snow-covered.

Downwelling radiation and wetness measurements may be comparable, remembering that there are likely to be climatologically driven differences. No other measurements can be reliably compared. Comparisons cannot be reliably made of the measurements of upwelling radiometer and surface soil heat flux

measurements at SGP site E21 (forest) versus other SEBS locations in SGP, because the surface vegetation types are so different; such comparisons are likely to show significant differences.

**Comparison of Data with the EBBR:** The only collocated EBBR and SEBS are at E13/E14 (SGP Central Facility). Caution must be used in the comparison of the two systems because they see different vegetation surfaces (Cook et al. 2006). The net radiation and surface soil heat flux measurements from the two will probably not be similar unless the ground is snow-covered.

**Comparison of Data with MET Instrumentation:** MET systems are collocated with ECOR/SEBS systems at most of the ECOR sites (exceptions are SGP E10 and E16). However, there are no measurements of the two systems that can be directly compared. SEBS net radiometer temperature (the temperature of the body of the net radiometer) is normally expected to be higher than MET temperature; they can be expected to be considerably different, especially under warm weather conditions.

**Comparison of Data with the ECOR:** No measurements of the two systems can be directly compared. SEBS net radiometer temperature (the temperature of the body of the net radiometer) is normally expected to be higher than ECOR sonic temperature but may be lower than the LI-7500 CO<sub>2</sub>/H<sub>2</sub>O analyzer temperature; they can be expected to be considerably different under warm weather conditions, especially.

**Common Conditions Reflecting Incorrect Data:** Upwelling and downwelling solar radiation measurements normally exhibit negative offsets, which do not correspond with physical reality. Measurements exceeding maximum QC values or less than minimum QC values can generally be interpreted as incorrect. Measurements that do not change at all over several hours (particularly if flat-lined) may not be correct. The following sources of information can be helpful in determining the value of the SEBS data.

See the QC flags in the SEBS data.

Conditions that commonly cause the SEBS downwelling radiation measurements to be incorrect include:

Periods of precipitation, fog, and dew (frost). This is caused by water lying on the upper domes of the net radiometer, thereby obstructing the passage of shortwave and longwave radiation (very light precipitation may have little or no effect). I do not write DQRs or indicate time periods in the monthly report for this wetting condition. The data user should look at the SEBS wetness measurement, the collocated or nearby MET rain gauges, or the DQ Explorer ECOR plots to determine times of precipitation and dew/frost. You can assume that offscale or spiked readings in the nighttime hours before dawn are normally caused by dew or frost on the upper net radiometer domes.

Large spikes (positive and negative) in `surface_energy_balance` and `surface_soil_heat_flux_avg` can occur when precipitation is heavy. The temperature of the water flowing into the soil from the soil surface (particularly during cold front passages), plus the effects of water/ice/snow on the net radiometer domes can cause such spikes.

## 6.0 Instrument Details

### 6.1 Detailed Description

#### 6.1.1 List of Components

**Net Radiometer:** CNR4/CNF4 by Kipp & Zonen

(<http://www.kippzonen.com/?product/85182/CNR+4.aspx>)

Spectral range: 300 to 2800 (shortwave) nm

Spectral range: 4500 to 42000 (longwave) nm

Sensitivity: 5 to 20  $\mu\text{V}/\text{W}/\text{m}^2$

Temperature dependence of sensitivity (-10 °C to +40 °C): < 4%

Response time: < 18 s

Non-linearity: < 1 %

Operating temperature: -40 to 80 °C

International standards (WMO): Good Quality WMO

Ventilation power: 10 W

**Wetness:** DRD11A Rain Detector by Vaisala, Inc.

(<http://www.vaisala.com/en/products/rainandprecipitationsensors/Pages/DRD11A.aspx>)

Range: 1 to 3 V DC (3 V dry, 1 V wet)

**Soil Heat Flow:** HFT-3 by Radiation and Energy Balance Systems, Inc.

Range: +/- 10 mV

**Soil Moisture:** SMP1 by Radiation and Energy Balance Systems, Inc.

capacitive sensor measuring resistance ratio

**Soil Temperature:** STP-1 by Radiation and Energy Balance Systems, Inc.

100-ohm platinum resistance temperature detector (PRTD)

**Data Logger:** CR1000 by Campbell Scientific, Inc.

**Multiplexers (2):** AM16/32B by Campbell Scientific, Inc.

#### 6.1.2 System Configuration and Measurement Methods

In a typical arrangement, the ECOR/SEBS system is placed on the north side of a crop field; the net radiometer is attached to the end of the ECOR boom, beneath the mount that holds the sonic and CO<sub>2</sub>/H<sub>2</sub>O sensor heads, at approximately 3 m above ground level (15 m height at SGP site EF21 forest). The wetness sensor is mounted on the top of the boom, midway between the two ends of the boom. The soil sensors are buried in the soil under the sonic and CO<sub>2</sub>/H<sub>2</sub>O sensor heads. The data logging equipment is installed in an enclosure attached to the ECOR tower.

Thirty-minute average measurements are stored within the CR1000 data logger and retrieved by the ECOR computer. The SEBS timestamp is the end of the half hour. The CR1000 data files are transferred

to the site data system for ingest (conversion into the netCDF format and incorporation of QC flags) and shipment to the ARM Data Center.

### **6.1.3 Specifications**

See instrument specifications in Section 6.1.1.

## **6.2 Theory of Operation**

The surface energy balance is determined from the net radiometer and soil sensor measurements. Upwelling and downwelling solar and infrared radiation measurements are combined to determine net radiation. The net radiometer measurements are adjusted to compensate for the temperature of the net radiometer body. Soil measurements are performed by three sets of soil heat flow (5 cm depth), soil temperature (0–5 cm average), and soil moisture (centered at 2.5 cm) probes. Soil heat flow is adjusted for the effect of soil moisture above the soil heat flow plate. The storage of energy in the soil above the soil heat flow plate is determined from the change in soil temperature with time.

Measurements from the three sets of soil probes are combined to give an average soil surface heat flux. Under vegetation canopies (particularly tall vegetation), the soil surface heat flux probably underestimates the heat flux from the top of the canopy. Soil surface heat flux and net radiation are combined to produce an estimate of the surface energy balance.

## **6.3 Calibration**

### **6.3.1 Theory**

All sensors are factory calibrated. Periodic in situ checks of the net radiometer calibrations are possible, but are not normally needed for years.

### **6.3.2 Procedures**

All sensors would be returned to the vendor for factory calibration, if needed.

### **6.3.3 History**

Not applicable.

## **6.4 Operation and Maintenance**

### **6.4.1 User Manual**

No single comprehensive user manual for the SEBS system is available for general use; rather, vendor-supplied documentation on sensors and a collection of procedures prepared by the mentor are provided for internal use by Site Operations.

## 6.4.2 Additional Documentation

Not applicable.

## 7.0 Glossary

See the ARM Glossary at <http://www.arm.gov/about/glossary>.

## 8.0 Citable References

Cook, DR, ML Fischer, and DJ Holdridge. 2006. “Comparison of ECOR, EBBR, and CO2FLX System Fluxes.” Sixteenth ARM Science Team Meeting, Albuquerque, New Mexico.

Kyrouac, J, and Y Hamada. 2018. “Comparison between co-located soil moisture measurements at the ARM Southern Great Plains (SGP) Central Facility.” Poster, ARM/ASR PI Meeting, Vienna, Virginia.

Webb, EK, GI Pearman, and R Leuning. 1980. “Correction of flux measurements for density effects due to heat and water vapour transfer.” *Quarterly Journal of the Royal Meteorology Society* 106(44): 85–100, <https://doi.org/10.1002/qj.49710644707>

